



# DIAGNOSTIC CARDIOLOGY SUITE WITH SPIROMETRY



SPECIFICATION	DESCRIPTION
Test Types Supported	FVC, SCV, MVV, Pre- and Post-bronchodilation
Sensor Type	Pneumotach
Accuracy & Reproducibility	Compliant with ATS/ERS 2005 Guidelines and ISO 26782 Guidelines
Flow Range	+/- 14 L/sec
Volume Range	0-14 L
Resistance	Compliant with ATS/ERS 2005 Guideline and ISO 26782 Guidelines (Impedance does not exceed 0.15kPa/ (L/s) with flows up to 14 L/s)
Interpretation	1991 ATS Interpretation Standards, Lung Age calculation available, Automatic Interpretation available, User-defined interpretation statements available
Parameters	FVC, FIVC, FIV1, FIV1%, FEV0.5, FEV1, FEV2, FEV3, FEV5, FEV6, FEV0.5, FEV0.5%, FEV1%, FEV1/FVC, FEV2%, FEV3%, FEV5%, FEV6%, PEF, FEF25, FEF50, FEF75, FEF0.2-1.2, FEF25-75, FEF75-85, PIF, FIF50, FEF50/FIF50, FEV1/FEV6, FET, MEF25, MEF50, MEF75 SVC, ERV, IRV, VT, IC, BF, MV, Tin, Tex, Tin/Tex MVV, MV, VT, BF, DFRC
Quality Checks	ATS Acceptability and ATS Reproducibility Effort-quality messages Audio and visual incentive for coaching patients

## MINIMUM SERVER REQUIREMENTS

SPECIFICATION	DESCRIPTION
Processor	Performance equivalent to an <b>Intel® Xeon®</b> processor class, quad-core with hyperthreading
Resolution	1024 x 768 or better
RAM	Minimum 4 GB
Operating System	<b>Microsoft</b> Server 2016 <b>Microsoft</b> Server 2019
Hard Drive Capacity	Minimum 250 GB (If client: minimum 160 GB)
System Disk	100 GB for OS and product installation (RAID recommended for data redundancy)
Data Disks	550 GB hard drive space available HD controller with 128 MB read/write cache (RAID recommended for data redundancy)

## MINIMUM STANDALONE/CLIENT SPECIFICATIONS

SPECIFICATION	DESCRIPTION
Operating Systems	<b>Microsoft Windows</b> 10 Professional and Enterprise (64-bit) <b>Microsoft Windows</b> 11 Professional and Enterprise Note: If installing <b>ELI Link</b> with <b>Diagnostic Cardiology Suite</b> system, requires 64-bit OS to operate both.
Compatible Viewing Resolutions and Font Scaling	<b>Resolutions:</b> 1366×768, 1680×1050, 1920×1080, 1920×1200, 2736×1824 <b>Font Scaling:</b> 100%
Processor	Performance equivalent or better than an Intel Core i3
RAM	Minimum 4 GB
Hard Drive Capacity	Minimum 250 GB (If client: minimum 160 GB)
Archive	Network or external USB drive
Software Installation	USB
Network	100 Mbps connection or better
USB Ports	Requires at least two USB 2.0 ports (spirometer external USB drive)
Input Devices	Standard keyboard, mouse, spirometer

## MINIMUM CITRIX® XENAPP® REQUIREMENTS

SPECIFICATION	DESCRIPTION
Client Machines That Run Citrix® XenApp®	<b>Microsoft Windows</b> 10 Professional and Enterprise (64-bit) <b>Microsoft Windows</b> 11 Professional and Enterprise <b>Citrix® Receiver</b> Any internet browser supported by <b>Citrix®</b>
Citrix® Desktop Delivery Controller	<b>Citrix® XenApp®</b> version 7.13 Any operating system supported by <b>Citrix®</b>
Citrix® App Servers	<b>Microsoft Server</b> 2016 <b>Microsoft Server</b> 2019 <b>Citrix® Virtual Delivery Agent</b> 7.13

## MINIMUM WINDOWS RDP REQUIREMENTS

SPECIFICATION	DESCRIPTION
Thin Clients Using Windows RDP	<b>Microsoft Windows</b> 10 Professional and Enterprise (64-bit) <b>Microsoft Windows</b> 11 Professional and Enterprise
Operating Systems for Servers Running Windows RDP	<b>Microsoft Server</b> 2016 <b>Microsoft Server</b> 2019

For more information, please contact your Baxter representative or call Baxter Customer Care on 1800 445 576.

**ALWAYS READ THE LABEL AND FOLLOW THE DIRECTIONS FOR USE**

CONTACT YOUR BAXTER REPRESENTATIVE FOR THE LATEST VERSIONS

Baxter reserves the right to make changes without notice in design, specifications and models.

Baxter, Diagnostic Cardiology Suite and ELI are trademarks of Baxter International Inc. or its subsidiaries.

Citrix and other marks appearing herein are trademarks of Citrix Systems, Inc., or its subsidiaries.

Intel and Xeon are trademarks of Intel Corporation or its subsidiaries.

Microsoft and Windows are trademarks of Microsoft Corporation.

AUTHORISED AUSTRALIAN SPONSOR Welch Allyn Australia Pty. Limited. 1 Baxter Drive, Old Toongabbie NSW 2146.

[hillrom.com.au](http://hillrom.com.au)

**Baxter**

ANZ-FLC51-240001\_r1\_04/2024